

## **X-Ray Area Posting Procedures**

Purpose: To comply with Arizona Administrative Code R-12-149 – Posting of Radiation Areas.

1. Prior to energizing a x-ray tube, all exterior doors of the radiation area will be posted with NAU EHS “X-Ray in Use Sign”.
2. All exterior doors of the radiation area must be locked when the x-ray tube is in operation.
3. NAU EHS “X-Ray in Use Sign” must be removed from exterior doors when procedures are completed and the x-ray tube is no longer energized.